

S E R V I C E N O T E

SUPERSEDES: None

37717A PDH/SDH/ATM Test Set**Serial Numbers:** Build Status 1:17 and below**Enhanced Jitter Measurement Accuracy****Situation:**

Improvements have been made to the PDH and SDH Jitter measurement accuracy of the 37717A by significantly reducing the instrument's intrinsics. The Jitter specifications on the latest data sheet shows these improvements.

Solution/Action:

Order Retrofit kit 15777A #026, this provides new firmware and hardware, and when fitted provides Enhanced Jitter Measurement Accuracy.

DATE: May 1995

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

DBG

ENTITY:

1400

ADDITIONAL INFORMATION: