

**53132A-09**

**S E R V I C E N O T E**

---

Supersedes:  
NONE

53132A- Universal Counter, 225 MHz, 12 digit/s, 150 ps.  
GPIB, RS232

Serial Numbers: ALL

Factory Cal Interval changed from 12 to 24 months effective 01-SEP-2011.

**Parts Required:**

P/N	Description	Qty.
NONE		

**ADMINISTRATIVE INFORMATION**

SERVICE NOTE CLASSIFICATION:	
<b>INFORMATION ONLY</b>	
<input type="checkbox"/> Calibration Required	PRODUCT LINE: SP
<input checked="" type="checkbox"/> Calibration NOT Required	AUTHOR: JD
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2013  
PRINTED IN U.S.A.

March 11, 2013

Rev. 21



**Situation:**

After extensive analysis of data gathered over a seven-year period, Agilent Technologies has extended the calibration cycle of its 53132A frequency counters from one year to two years.

Product stability and aging rates, and all performance specifications provided in the data sheet, remain unchanged and guaranteed with the two-year calibration cycle. There will be no changes to the product documentation (data sheet, manuals, web content) that mention the change to the two-year calibration cycle.

**Solution/Action:**

NONE

**Revision History:**

Revision Number	Date	Author	Reason For Change
1.0	3/7/13	JD	As published