

S E R V I C E N O T E

SUPERSEDES: None

E4916A Crystal Impedance Meter**Serial Numbers:** JP1KD00000 / JP1KD00391**Replacement for A1 Main Board.****Duplicate Service Notes:**

E4915A-02 Serial Numbers: JP1KD00000 / JP1KD00128

To Be Performed By: Agilent-Qualified Personnel**Situation:**

The A1 main board has been obsolete as discontinuance of some parts. Then we released the new A1 board as replace for the old A1 board.

Solution / Action:

When the A1 board is faulty, order the either of following parts.

P/N:E4915-66511 A1 main board

P/N:E4915-69511 A1 main board (re-built board of the above)

Note:

If the E4916A is F/W rev 2.01 or smaller, order the following kit.

P/N:E4915-18005 Firmware Upgrade Kit (Rev 2.02) Because F/W rev 2.01 or smaller only meet the old A1 board, but this F/W rev 2.02 meet both of the new A1 and the old A1 board,

DATE: December 1997

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATI ON ONLY

AUTHOR:

ENTITY:

ADDITIONAL INFORMATION:

HM

3355

