

INFORMATION ONLY – DOES NOT COMMUNICATE
A MODIFICATION OR SAFETY CONDITION

E6432A-01

S E R V I C E N O T E

Supersedes:
None

E6432A VXI Microwave Synthesizer, 10MHz to 20GHz

Serial Numbers: [0000A00000 / 9999Z99999]

E6432A calibration interval extended from 12 months to 36 months

Parts Required: None

Situation:

Based on the performance and design of the E6432A, the calibration cycle has been extended from 12 months to 36 months. This change is effective immediately and applies to all units that are under current calibration.

Solution/Action:

Customers with E6432A may elect to extend the calibration of their instrument if it currently has a valid calibration.

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR: MHF PRODUCT LINE: BL

ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2004
PRINTED IN U.S.A.



Agilent Technologies

January 12, 2004