

E8257D-25

S E R V I C E N O T E

Supersedes:
NONE

E8257D Analog Signal Generator

Serial Numbers: YY49280000- YY49289999

TME software support's delayed for E8257D's with serial number prefix YY4928XXXX.

Parts Required:

P/N	Description	Qty.
-----	-------------	------

NONE

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
INFORMATION ONLY	
AUTHOR: SN	PRODUCT LINE: 15
ADDITIONAL INFORMATION: This service note will be expired in May, 2010	

© AGILENT TECHNOLOGIES, INC. 2008
PRINTED IN U.S.A.



August 3, 2009

Situation:

The E8257D with serial prefix number YY4928XXXX has improved phase noise and higher power specifications. The Automatic Leveling Control (A10) and the Low band output board (A8) have been modified. Also the Amplifier Filter Assembly, A38 has been added.

TME Calibration Software needs to be modified for these hardware changes and specifications.

- Performance Verification Software will be released December 21, 2009.
- Adjustments Software will be released May 10, 2010.

Solution/Action:

1) Z540 and 17025 for E8257D with serial number prefix YY4928XXXX will not be orderable, from the factory or service center until December 21, 2009

- Note: For E8257D's with serial number prefix < YY4928XXXX, the Z540 and 17025 can still be tested with the current TME software.

2) Repair and calibration for E8257D with serial prefix YY4928XXXX, is return-to-factory until May 10, 2010 TME software release.