

E8363C-03

S E R V I C E N O T E

Supersedes:
NONE

E8363C PNA Microwave Network Analyzers

Serial Numbers: US000000000/US999999999 (ALL)

During System Verification test, E8363C analyzers that encounter errors while measuring the verification kit devices should upgrade to the most recent firmware and retry the test.

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
INFORMATION ONLY	
AUTHOR: MF	PRODUCT LINE: WN
ADDITIONAL INFORMATION: Affected PNA models: E8361/2/3/4C and N5230C. Service Note history: E8363C-03 written August 31, 2009.	

© AGILENT TECHNOLOGIES, INC. 2009
PRINTED IN U.S.A.

September 1, 2009

Rev. 16



Agilent Technologies

Situation:

During System Verification test, some E8363C analyzers are experiencing errors while measuring the verification kit devices.

Solution/Action:

Download the most recent firmware and retry the test.