

# N5242A-01C

## S E R V I C E N O T E

Supersedes:  
N5242A-01B

### N5242A PNA-X Microwave Network Analyzers

**Serial Numbers:** AA00000000/ZZ99999999 (ALL)

N5242A PNA-X Network analyzers have RESUMED shipping with front handles.

**Parts Required:**

P/N	Description	Qty.
None	--	--

### ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
<b>INFORMATION ONLY</b>	
AUTHOR: PYL	PRODUCT LINE: WN
ADDITIONAL INFORMATION: Affected PNA models: N5242A. Service Note History: N5242A-01 released 1/14/08; N5242A-01A released 1/23/08: removed handle p/n; N5242A-01B: released 1/31/08: added last 2 paragraphs; N5242A-01C: released 10/17/08: info revised because front handles shipping again.	

© AGILENT TECHNOLOGIES, INC. 2008  
PRINTED IN U.S.A.

October 28, 2008

Rev. 15



**Situation:**

Agilent Technologies has completed the testing of PNA-X front handles. The testing concluded that each front handle is designed to carry 4-times the weight of a PNA-X unit. Safety load tests and other necessary parameter tests have been qualified and validated.

**Solution/Action:**

Front handles have resumed shipping with the PNA-X, and customers can again order the option N5242A-1CP rack mount kit. If you have any questions, please contact Agilent using the information below.

Sincerely,  
Agilent Technologies  
Component Test Division Customer Support  
(707) 577- 6802  
[CTD-SOCO\\_SUPPORT@agilent.com](mailto:CTD-SOCO_SUPPORT@agilent.com)